Revision No.: ZJ0007
Issue Date: Jul 1st, 2012

Japanese Encephalitis Virus (JEV) Real Time RT-PCR Kit User Manual

For In Vitro Diagnostic Use Only

REF **ER-0065-01**

For use with LightCycler1.0/2.0 Instrument

EC REP

1. Intended Use

The Japanese encephalitis virus real time RT-PCR kit is used for the detection of Japanese encephalitis virus in serum, plasma, C.S.F or mosquito sample by using real time PCR systems.

2. Principle of Real-Time PCR

The principle of the real-time detection is based on the fluorogenic 5'nuclease assay. During the PCR reaction, the DNA polymerase cleaves the probe at the 5' end and separates the reporter dye from the quencher dye only when the probe hybridizes to the target DNA. This cleavage results in the fluorescent signal generated by the cleaved reporter dye, which is monitored real-time by the PCR detection system. The PCR cycle at which an increase in the fluorescence signal is detected initially is proportional to the amount of the specific PCR product. Monitoring the fluorescence intensities in real-time allows the detection of the accumulating product without having to re-open the reaction tube after the amplification.

3. Product Description

Japanese encephalitis is a mosquito-borne disease caused by the Japanese encephalitis virus. Over 50,000 cases are reported to occur each year. The disease is found in over 25 countries throughout the world. It occurs mainly in the rural and agricultural areas of Asia and Western Pacific. In temperate regions such as China, Japan and Korea, the transmission is highest from April to September. The disease is transmitted through the bite of an infected female culex mosquito. In the majority of cases these may occur within rural areas of the endemic countries and especially in regions where pig farming is found. The incubation period is usually five to 15 days. The majority of Japanese B encephalitis infections are subclinical (have no symptoms) and many infected people become symptomless carriers. Mild infections occur with few apparent symptoms other than fever with headache. Acute encephalitis can progress to paralysis, seizures, coma and death.

The Japanese encephalitis virus real time RT-PCR kit contains a specific ready-to-use system for the detection of the Japanese encephalitis virus using RT-PCR (reverse transcription polymerase chain reaction) in the real-time PCR system. The master contains a Super Mix for the specific amplification of the Japanese encephalitis virus RNA. The reaction is done in one step real time RT-PCR. The first step is a reverse transcription (RT), during which the Japanese encephalitis virus RNA is transcribed into cDNA. Afterwards, a thermostable DNA polymerase is used to amplify the specific gene fragments by means of PCR (polymerase chain reaction). Fluorescence is emitted and measured by the real time systems´ optical unit during the PCR. The detection of amplified Japanese encephalitis virus DNA fragment is performed in fluorimeter **channel 530nm** with the fluorescent quencher BHQ1. In addition, the kit contains a system to identify possible PCR inhibition by measuring the HEX/VIC/JOE fluorescence of the internal control (IC).

4. Kit Contents

Ref.	Type of reagent	Presentation 25rxns
1	JEV Super Mix	1 vial, 350µl
2	RT-PCR Enzyme Mix	1 vial, 28µl
3	Molecular Grade Water	1 vial, 400µl
4	Internal Control (IC)	1 vial, 30µl
5	JEV Positive Control	1 vial, 30µl

Analysis sensitivity: 1×10^4 copies/ml

Note: Analysis sensitivity depends on the sample volume, elution volume, nucleic acid extraction methods and other factors .If you use the RNA extraction kits recommended, the analysis sensitivity is the same as it declares. However, when the sample volume is dozens or even hundreds of times greater than elution volume by some concentrating method, it can be much higher.

5. Storage

- All reagents should be stored at -20°C. Storage at +4°C is not recommended.
- All reagents can be used until the expiration date indicated on the kit label.
- Repeated thawing and freezing (> 3x) should be avoided, as this may reduce the sensitivity of the assay.
- Cool all reagents during the working steps.
- Super Mix should be stored in the dark.

6. Additionally Required Materials and Devices

- Biological cabinet
- Real time PCR system
- Desktop microcentrifuge for "eppendorf" type tubes (RCF max. 16,000 x g)
- Vortex mixer
- RNA extraction kit
- Real time PCR reaction tubes/plates
- Cryo-container
- Pipets (0.5 µl – 1000 µl)
- Sterile filter tips for micro pipets
- Sterile microtubes
- Disposable gloves, powderless
- Biohazard waste container
- Refrigerator and freezer
- Tube racks

7. ⚠Warnings and Precaution

- Carefully read this instruction before starting the procedure.
- For in vitro diagnostic use only.
- This assay needs to be carried out by skilled personnel.
- Clinical samples should be regarded as potentially infectious materials and should be prepared in a laminar flow hood.
- This assay needs to be run according to Good Laboratory Practice.
- Do not use the kit after its expiration date.
- Avoid repeated thawing and freezing of the reagents, this may reduce the sensitivity of the test.
- Once the reagents have been thawed, vortex and centrifuge briefly the tubes before use.
- Prepare quickly the Reaction mix on ice or in the cooling block.
- Set up two separate working areas: 1) Isolation of the RNA/ DNA and 2) Amplification/ detection of amplification products.
- Pipets, vials and other working materials should not circulate among working units.
- Use always sterile pipette tips with filters.
- Wear separate coats and gloves in each area.
- Do not pipette by mouth. Do not eat, drink, smoke in laboratory.
- Avoid aerosols.

8. Sample Collection, Storage and transport

- Collected samples in sterile tubes.
- Specimens can be extracted immediately or frozen at -20°C to -80°C.
- Transportation of clinical specimens must comply with local regulations for the transport of etiologic agents.

9. Procedure

9.1 RNA-Extraction

RNA extraction kits are available from various manufacturers. You may use your own extraction systems or the commercial kit based on the yield. For the RNA extraction, please comply with the manufacturer's instructions. The recommended extraction kit is as follows:

Nucleic Acid Isolation Kit	Cat. Number	Manufacturer
RNA Isolation Kit	ME-0010/ME-0012	ZJ Biotech
QIAamp Viral RNA Mini extraction Kit (50)	52904	QIAGEN

9.2 Internal Control

It is necessary to add internal control (IC) in the reaction mix. Internal control (IC) allows the user to determine and control the possibility of PCR inhibition.

Add the internal control (IC) 1µl/rxn and the result will be shown in the 560nm.

9.3 RT-PCR Protocol

The Master Mix volume for each reaction should be pipetted as follows:



※PCR system without 560nm channel may be treated with 1µl Molecular Grade Water instead of 1µl IC.

1) The volumes of Super Mix and Enzyme Mix per reaction multiply with the number of samples, which includes the number of controls, and sample prepared. Molecular Grade Water is used as the negative control. For reasons of unprecise pipetting, always add an extra virtual sample. Mix completely then spin down briefly in a centrifuge.
2) Pipet **15µl** Master Mix with micropipets of sterile filter tips to each of the real time PCR reaction plate/tubes. Separately add **5µl** RNA sample, positive and negative controls to different reaction plate/tubes. Immediately close the plate/tubes to avoid contamination.
3) Spin down briefly in order to collect the Master Mix in the bottom of the reaction tubes.
4) Perform the following protocol in the instrument:

45°C for 10min	1cycle
95°C for 15min	1cycle
95°C for 5sec, 60°C for 30sec (Fluorescence measured at 60°C)	40cycles

Selection of fluorescence channels	
530nm	Target Nucleic Acid
560nm	IC

10. Threshold setting: Choose **Arithmetic** as back ground and **none** as Noise Band method, then adjust the Noise band just above the maximum level of molecular grade water, and adjust the threshold just under the minimum of the positive control.

11. Quality control: Negative control, positive control and internal control must be performed correctly, otherwise the sample results is invalid.

Channel / Control	Crossing point value	
	530nm	560nm
Molecular Grade Water	Blank	25~35
Positive Control(qualitative assay)	≤35	——

12. Data Analysis and Interpretation

The following results are possible:

	Crossing point value		Result Analysis
	530nm	560nm	
1#	Blank	25~35	Below the detection limit or negative
2#	≤38	——	Positive
3#	38~40	25~35	Re-test; if it is still 38~40, report as 1#
4#	Blank	Blank	PCR Inhibition; no diagnosis can be concluded.

Gentaur Molecular Products
Voortstraat 49
1910 Kampenhout, BELGIUM

Tel 0032 16 58 90 45 | Fax 0032 16 50 90 45
www.gentaur-worldwide.com
info@gentaur.com